

Documentation

Explanatory Notes

The IT'IS materials parameter database aims to provide the most complete, reliable and up-to-date database of EM and thermal tissue parameters and their uncertainty. Particular care is taken to provide material parameters for all of the tissues included in the 'Virtual Population' models. The thermal and dielectric parameters found in the database are calculated based on a critical study of the existing literature. Large variations in reported values exist for some parameters, which are due to factors such as the method used, the sample size and physiological conditions of studied subjects, intersubject variations, measurement conditions, etc. In determining which studies to use to calculate our values, we had to decide whether to be very selective and use only the studies that we judged to be the most accurate, or to be more inclusive and incorporate studies that may have varying degrees of accuracy. We decided on the latter approach (after eliminating studies with major flaws) in order to increase the parameter sample size used to determine the average and to gain information about the variability of the parameters. General considerations used while establishing the database are detailed below.

The thermal parameter values were calculated for five key properties: tissue perfusion rate (heat transfer rate, HTR), thermal conductivity (TC), heat capacity (HC), metabolic heat production (heat generation rate, HGR), and density. A large number of different studies can be found throughout the literature for the different parameters, in particular for the blood perfusion rate. Data on these five properties are represented unevenly in the scientific literature and many educated judgments were made in order to arrive at the best values.

Determining the HTR presented a considerable challenge. One reason for this is the natural variability within the species caused by the current physiological condition of the studied subject. In order to compensate for the different values, we calculated the mean HTR (ml/min per kg of tissue) from data compiled from studies performed in humans and animals. If values were recorded as percent of cardiac output (% CO), we calculated the mean blood flow rate based on the standard conditions proposed by Williams & Leggett [1]. In most cases we included only data from young and healthy subjects. When data from young subjects was unavailable, such as values for tongue tissue, we included data from older subjects.

Another problem with determining the HTR is the lack of published material for many organs. Data from animal studies were included when no human data was available. Where possible, we included only animals that have similar physiological properties to humans and should therefore also have similar tissue perfusion rates. The methodology for animal experiments was at times more accurate than for human experiments. In these cases we included the values from the human as well as the animal studies in our calculations. When no experimental data was available for a given tissue, we determined the HTR based on that of tissues of similar function and/or tissue composition.

TC and HC, two of the key properties in thermal parameters, are strongly dependent on the composition of the tissue, in particular on the tissue water content. TC and HC measurements can only be found in the literature for a selected number of tissues. When no data was available, we calculated the TC and HC based on the tissue water content as proposed by McIntosh and Anderson [2]. If the water content could also not be specified, we assumed TC and HC to be similar to an organ of similar tissue composition and/or function. If a tissue is a mixture of two tissues with known TC and HC, we calculated the average of the two tissues (e.g. thymus is the average of lymph node and fat).

As we also used review articles for the determination of TC and HC, values from the same original study may have entered our calculations several times. However, this should not have a strong influence on the calculated mean as most values are within a narrow range. Nevertheless, this issue will be resolved in future versions of our database.

The HGR is proportional to the HTR and was calculated based on the formula proposed by McIntosh and Anderson [2], with some modifications due to what appears to be unit conversion errors in the original formula: to convert from cal/100g/min to W/kg we used a factor of 0.6973 instead of 0.4184 as used by McIntosh and Anderson. The HTR values used in the formula were derived for the IT'IS materials database as explained above.

The density values were derived from several publications. Where no values could be determined for a specific tissue, we assumed the density to be equal to a tissue of similar composition and/or function. Similar to TC and HC, if a tissue is a mixture of two tissues with known densities we calculated the average of the two tissues.

Our dielectric parameter values are based on the tissue dielectric property database generated by Gabriel et al. [3]. The dielectric properties for this database were calculated for a frequency spectrum ranging from single Hz to several GHz. As this spectrum contains four dispersion regions, the values can be fitted using a 4-cole-cole dispersion model. However, the Gabriel et al. database contains only a restricted number of organs and tissues. For tissues not included in work presented by Gabriel et al., we took the dielectric values of an organ of similar function and/or tissue composition with known dielectric properties. More work has been published about measurements of dielectric properties of biological tissues, although most of them focus on a specific frequency range. As these cannot be described by a 4-cole-cole expression, they were not included in this database. Some applications in specific frequency ranges could gain from those values. Users are encouraged to discuss these possibilities in our forum: forum.database.itis.ethz.ch.

References

1. L. Williams and R. Leggett. Reference values for resting blood flow to organs of man. *Clinical Physics and Physiological Measurement*, 10: 187-217, 1989.
2. R. L. McIntosh and V. Anderson, A Comprehensive Tissue Properties Database Provided for the Thermal Assessment of a Human at Rest. *Biophysical Reviews and Letters*, 05(03): 129-151, 2010.
3. C. Gabriel. Compilation of the Dielectric Properties of Body Tissues at RF and Microwave Frequencies, Report N.AL/OE-TR- 1996-0037, Occupational and environmental health directorate, Radiofrequency Radiation Division, Brooks Air Force Base, Texas (USA), 1996.

Known Issues

The following list contains some of the issues that arose during the compilation of the material parameter database. Whenever possible, we will address these issues in future updates.

- There is no data available in the literature for the dielectric properties of urine over the frequency spectrum (single Hz to a few GHz) that is mandatory for a 4 cole-cole fit. Therefore, for the dielectric parameter values of urine, we have chosen the values for the urinary bladder wall and for the thermal properties we calculated the mean value of bladder wall and urine.
- The dielectric properties database generated by Gabriel et al. [1] contains only values for a selected number of glands (thyroid, testes and ovary). For all glands of the reproductive system we choose the same values as for testes and for all other glands the same values as for thyroid gland.
- It is a well documented phenomenon that the thymus increases in size until puberty and then, with advancing age, undergoes a process known as involution where the thymus slowly reduces in size and is replaced by adipose tissue [2]. Since the percentage of replaced thymus tissue varies with age, we have chosen for the thermal properties the mean values between thymus and fat tissue. However, since in young adults most of the thymus did not yet turn into fat tissue but in elderly people the majority of the thymus underwent already involution, a different ratio between thymus and fat may be calculated. For the dielectric properties we have chosen to take the 4-cole-cole fit corresponding to lymph node. However, based on the age of the model in use, fat may be selected for the calculation of the dielectric properties.
- The content of the stomach and intestine depends on the diet of the subject. For the thermal properties we calculated an average of water and muscle which would correspond to a diet of 50% of water and 50% of meat. For the dielectric properties we have chosen the values for muscle only.

References

1. C. Gabriel. Compilation of the Dielectric Properties of Body Tissues at RF and Microwave Frequencies, Report N.AL/OE-TR- 1996-0037, Occupational and environmental health directorate, Radiofrequency Radiation Division, Brooks Air Force Base, Texas (USA), 1996.
2. M. Nishino, S. K. Ashiku, O. N. Kocher, R. L. Thurer, P. M. Boiselle and H. Hatabu. The Thymus: a Comprehensive Review. *Radiographics*, 26:335, 2006.